

Using healthPAL™ with the **NONIN 9560 Pulse Oximeter**

From the MedApps HealthPAL QuickStart Guide

Collecting Blood Oxygen Saturation

1. Ensure that Pulse Oximeter is powered OFF.
2. Squeeze the Pulse Ox at the bottom of device to “open” – then insert your finger, nail side up, until fingertip touches the built-in “stop” guide.
3. Pulse Oximeter will automatically power ON.
4. Pulse Oximeter’s LEDs will blink (approximately 5 seconds) as it assesses SpO2 and pulse rate. The Pulse Ox will display the measurement.
5. HealthPAL will display and audibly announce the reading.
(**“Your pulse ox reading is XX %.”**)

The measurement will be stored in HealthPAL’s memory and transmitted to server automatically when HealthPAL is inactive for 3 minutes.

Always follow manufacturer instructions for operation and safety for any accessory health monitors used with HealthPAL.

